

## Contents

Where Movement is Provided

What Happens if Movement is not Provided?

Dynamic Solutions

Expected Outcomes

## Released

October 2017

## Contributors

K. Missy Ball, MT, PT, ATP

Michelle L. Lange, OTR/L, ABDA, ATP/SMS

Sharon Sutherland, PT

## Editors

Michelle Lange

Greg Peek

## Prepared by

Seating Dynamics

## Movement and Behavioral Issues

### Where Movement Occurs, What Happens if Movement is not Provided, Dynamic Solutions, and Expected Outcomes

#### Introduction

Dynamic seating provides movement within a wheelchair. When the client moves, the dynamic seating components move with the client, maintaining alignment with the seating system for postural support, skin integrity preservation, and stability. Clients who seek out movement may benefit from dynamic seating. These clients may simply choose to move, stretch, or even rock repeatedly. Dynamic components absorb and diffuse force, protecting both the client and the wheelchair seat and frame from damage. Providing movement can increase alertness, decrease agitation, and reduce maladaptive behaviors.

Clients with a High Threshold sensory processing modulation disorder seek out sensory input, secondary to the inability to detect where their head is in space in relation to gravity within normal sensory parameters. Banging against the back or bouncing on the seat provides increased tactile and vestibular sensory input, providing the client with feedback necessary to calm and quiet their emotional as well as their physical state for postural improvement. Dynamic seating can improve quality of life for many wheelchair users and their caregivers.

This document is designed to provide Clinical Indicators for the use of Dynamic Seating with clients who seek out movement and /or have behavioral issues. Please refer to our other Clinical Indicators for additional Dynamic Seating applications.

#### Where Movement Occurs: Hips / Knees / Ankles

##### What Happens if Movement is not Provided?

- 1) The client may continue to rock or bang against unyielding surfaces. This could lead to the wheelchair moving or tipping over.
- 2) Breakage can occur due to strong and repeated movements at:
  - a) Footrest hangers and footplates
  - b) Back canes
  - c) Seat frame
  - d) Seating system mounting hardware
  - e) Head support mounting hardware
- 3) Breakage can occur while in transport, placing the client at risk of injury.
- 4) Injury can occur:
  - a) Client may be injured from the wheelchair tipping over or moving.
  - b) Client may be injured from significant and repeated forces against the seating system and mobility base.
  - c) Client may be injured from contact with sharp, broken objects after or during breakage.

- 5) Loss of desired or optimal postural alignment can occur:
  - a) Client may move out of a beneficial position, resulting in poor alignment, poor pressure distribution, and decreased function.
  - b) Poor alignment may specifically reduce trunk and head alignment and control.
- 6) Pain or discomfort can occur:
  - a) Client may experience discomfort / pain due to significant and repeated forces against the seating system and mobility base.
  - b) Pain or discomfort can decrease sitting tolerance.
- 7) Excessive energy expenditure can occur as the client continues to move within the static seating system and mobility base. This can increase caloric output, body temperature, sweating, and fatigue.
- 8) Client agitation may occur if movement is not allowed. Movement can promote calming.
- 9) Client alertness may decrease if movement is not allowed. Movement can increase alertness.
- 10) Maladaptive behaviors may be reduced by allowing movement within the seating system and mobility base. (Lange, et al, 2017; Lange, 2013; Presperin-Pedersen & Eason, 2015)

### **What Dynamic Components may be used to Address or Reduce the Likelihood of the Above Issues?**

- 1) Dynamic Back support hardware
- 2) Dynamic Legrest support hardware
- 3) Dynamic Head support hardware
- 4) Dynamic secondary support components (i.e. shoulder straps)

### **Expected Outcomes:**

- 1) When client movement occurs, the dynamic component will move, forces will be diffused (Avellis et al, 2010), and the energy built up in the dynamic component will return the client to their starting position without loss of postural alignment.
- 2) Due to force diffusion:
  - a) Reduced equipment breakage (Crane et al., 2007)
  - b) Reduced client injury (Lange et al., 2017)
  - c) Reduced loss of posture (Crane et al., 2007; Ferrari, 2003)
  - d) Reduced energy expenditure (Ferrari, 2003)
- 3) Due to provision of movement:
  - a) The seating components will move with the client during any intermittent or continual rocking patterns
  - b) Increased sitting tolerance / comfort (Crane et al., 2007)
  - c) Increased sensory input (Presperin-Pedersen & Eason, 2015)
  - d) Increased alertness (Pfeiffer, et al, 2008)
  - e) Reduced agitation (Watson, et al, 1998)
  - f) Reduced maladaptive behaviors (Pfeiffer, et al, 2011)

## Where Movement Occurs: Torso (Spinal Extension, similar characteristics as Hip Extension)

### What Happens if Movement is not Provided?

- 1) The client may continue to rock or bang against unyielding surfaces. This could lead to the wheelchair moving or tipping over.
  - 2) Breakage can occur due to strong and repeated movements at:
    - a) Footrest hangers and footplates
    - b) Back canes
    - c) Seat frame
    - d) Seating system mounting hardware
    - e) Head support mounting hardware
  - 3) Breakage can occur while in transport, placing the client at risk of injury.
  - 4) Injury can occur:
    - a) Client may be injured from the wheelchair tipping over or moving.
    - b) Client may be injured from significant and repeated forces against the seating system and mobility base.
    - c) Client may be injured from contact with sharp, broken objects after or during breakage.
  - 5) Loss of desired or optimal postural alignment can occur:
    - a) Client may move out of a beneficial position, resulting in poor alignment, poor pressure distribution, and decreased function.
    - b) Poor alignment may specifically reduce trunk and head alignment and control.
  - 6) Pain or discomfort can occur:
    - a) Client may experience discomfort / pain due to significant and repeated forces against the seating system and mobility base.
    - b) Pain or discomfort can decrease sitting tolerance.
  - 7) Excessive energy expenditure can occur as the client continues to move within the static seating system and mobility base. This can increase caloric output, body temperature, sweating, and fatigue.
  - 8) Client agitation may occur if movement is not allowed. Movement can promote calming.
  - 9) Client alertness may decrease if movement is not allowed. Movement can arouse and increase alertness.
  - 10) Maladaptive behaviors may be reduced by allowing movement within the seat and mobility base.
- (Lange, et al, 2017; Lange, 2013; Presperin-Pedersen & Eason, 2015)

### What Dynamic Components may be used to Address or Reduce the Likelihood of the Above Issues?

- 1) Dynamic Back support hardware
- 2) Dynamic Legrest support hardware
- 3) Dynamic Head support hardware
- 4) Dynamic secondary support components (i.e. shoulder straps)

## Expected Outcomes:

- 1) When client movement occurs, the dynamic component will move, forces will be diffused (Avellis et al, 2010), and the energy built up in the dynamic component will return the client to their starting position without loss of postural alignment.
- 2) Due to force diffusion:
  - a) Reduced equipment breakage (Crane et al., 2007)
  - b) Reduced client injury (Lange et al., 2017)
  - c) Reduced loss of posture (Crane et al., 2007; Ferrari, 2003)
  - d) Reduced energy expenditure (Ferrari, 2003)
- 3) Due to provision of movement:
  - a) The seating components will move with the client during any intermittent or continual rocking patterns
  - b) Increased sitting tolerance / comfort (Crane et al., 2007)
  - c) Increased sensory input (Presperin-Pedersen & Eason, 2015)
  - d) Increased alertness (Pfeiffer, et al, 2008)
  - e) Reduced agitation (Watson et al., 1998)
  - f) Reduced maladaptive behaviors (Pfeiffer, et al, 2011)

## Where Movement Occurs: Cervical

### What Happens if Movement is not Provided?

- 1) The client may continue to rock or bang against unyielding surfaces. This could lead to the wheelchair moving or tipping over. A central gravity axis (CG) tilt is recommended for clients needing a tilt who display excessive movement, to prevent destabilization of the wheelchair frame. Also the CG tilt is a very stable base on which to add dynamic components.
- 2) Breakage can occur at the head support mounting hardware. Even if breakage does not occur, excessive forces can move the head support out of alignment.
- 3) Breakage can occur while in transport, placing the client at risk of injury
- 4) Injury can occur:
  - a) Client may be injured from the wheelchair tipping over or moving.
  - b) Client may be injured from significant and repeated forces against the head support pads. A concussion can occur when the head collides with force against a surface. According to the Dr. Kim Gorgens of the University of Denver (Science Daily; June, 2010), a concussion can occur at 95 g's. G-Force is a unit of force equal to the force exerted by gravity. The average football player receives 103 g's when hit during a game. Some clients using wheelchairs bang against the head pad with significant force, perhaps even enough force to cause brain injury. Degree of force and repetitive impacts only increase risk of injury.
  - c) Client may be injured from contact with exposed hardware if head support pad and/or mount break.
- 5) Loss of desired or optimal postural alignment can occur:
  - a) Poor alignment may specifically reduce trunk and head alignment and control.
  - b) If the head moves into hyperextension as the result of undiffused forces or if the head support moves out of position, reflexive responses may be elicited and postural insecurity, startle, and anxiety may increase.

- c) The client may come off of the head pad which can lead to the client being caught under the head pad and/or choking.
- 6) Pain or discomfort can occur:
  - a) Client may experience discomfort / pain due to significant and repeated forces against the head support.
  - b) Pain or discomfort can decrease sitting tolerance
- 7) Excessive energy expenditure can occur as the client continues to move within the static seating system and mobility base. This can increase caloric output, body temperature, sweating, and fatigue.
- 8) Client agitation may occur if movement is not allowed. Movement can promote calming.
- 9) Client alertness may decrease if movement is not allowed. Movement can arouse and increase alertness.
- 10) Maladaptive behaviors may be reduced by allowing movement within the seat and mobility base. (Lange, et al, 2017; Lange, 2013; Presperin-Pedersen & Eason, 2015)

### **What Dynamic Components may be used to Address or Reduce the Likelihood of the Above Issues?**

- 1) Dynamic Back support hardware
- 2) Dynamic Head support hardware

### **Expected Outcomes:**

- 1) When client movement occurs, the dynamic component will move, forces will be diffused (Avellis et al., 2010), and the energy built up in the dynamic component will return the client to their starting position without loss of postural alignment.
- 2) Due to force diffusion:
  - a) Reduced equipment breakage (Crane et al., 2007)
  - b) Reduced client injury (Lange et al., 2017)
  - c) Reduced loss of posture (Crane et al., 2007; Ferrari, 2003)
  - d) Reduced energy expenditure (Ferrari, 2003)
- 3) Due to provision of movement:
  - a) The seating components will move with the client during any intermittent or continual rocking patterns
  - b) Increased sitting tolerance / comfort (Crane et al., 2007)
  - c) Increased sensory input (Presperin-Pedersen & Eason, 2015)
  - d) Increased alertness (Pfeiffer, et al, 2008)
  - e) Reduced agitation (Watson, et al, 1998)
  - f) Reduced maladaptive behaviors (Pfeiffer, et al, 2011)

## References:

1. Avellis, M., Cazzaniga, A., Cimolin, V., Galli, M., and Turconi, A.C. (2010). Dynamic seating vs. rigid seating: A quantitative comparison using 3d movement analysis in people with cerebral palsy. *Posture and Mobility*, 26(1):15–16.
2. Crane, B. A., Holm, M. B., Hobson, D., Cooper, R. A., & Reed, M. P. (2007). A dynamic seating intervention for wheelchair seating discomfort. *American Journal of Physical Medicine & Rehabilitation*, 86(12), 988-993.
3. Ferrari A. (2003). "In terms of posture and postural control", *Giornale Italiano di Medicina Riabilitativa*, 17 (1); 61-7.
4. Gorgens, K. (2010). "Most concussions deliver 95 g's, neuropsychologist says," *Science News*, Science Daily, June 25, 2010.
5. Lange, M. (2013). Dynamic Seating Webinar. National Seating and Mobility. Retrieved from <https://vimeo.com/100236246>.
6. Lange, M., Pedersen, J., Sparacio, J., Eason, S., Sutherland, S. (2017). Dynamic Seating –Providing Movement and Why. International Seating Symposium, Nashville, TN.
7. Pedersen, J. and Eason, S. (2015). Using Seating to Enhance Movement of the Body in the Wheelchair. International Seating Symposium, Nashville, TN.
8. Pfeiffer, B. A., Koenig, K., Kinnealey, M., Sheppard, M., & Henderson, L. (2011). Effectiveness of sensory integration interventions in children with autism spectrum disorders: A pilot study. *American Journal of Occupational Therapy*, 65(1), 76-85.
9. Pfeiffer, B., Henry, A., Miller, S., & Witherell, S. (2008). Effectiveness of Disc 'O'Sit cushions on attention to task in second-grade students with attention difficulties. *American Journal of Occupational Therapy*, 62(3), 274-281.
10. Watson, N. M., Wells, T. J., & Cox, C. (1998). Rocking chair therapy for dementia patients: Its effect on psychosocial well-being and balance. *American Journal of Alzheimer's Disease and Other Dementias*, 13(6), 296-308.

